

EU Declaration of Conformity

| | | | |
|---------------------|--|--------------------|---------------|
| Product | Laboratory Equipment | File Number | TFR920111_001 |
| Manufacturer | Cole-Parmer Ltd Beacon Road Stone, Staffordshire ST15 0SA United Kingdom | | |

This declaration of conformity is issued under the sole responsibility of the manufacturer

Object of Declaration

Flame Photometer Range
(reference the attached list of catalogue numbers)


The object of the declaration described above is in conformity with the relevant Union Harmonisation Legislation:

| | |
|------------------------------|-------------|
| Low Voltage Directive | 2006/95/EC |
| EMC Directive | 2004/108/EC |
| RoHS Directive | 2011/65/EC |

References to the relevant harmonised standards used or references to the other technical specifications in relation to which conformity is declared:

| | |
|-------------------------|--|
| IEC/EN 61010-1:2001 | Safety requirements for electrical equipment for measurement, control and laboratory use. Part 1: General requirements. |
| IEC/EN 61010-2-010:2003 | Particular requirements for laboratory equipment for the heating of materials. |
| IEC/EN 61326-1:2006 | Electrical equipment for measurement, control and laboratory use. EMC requirements. Part 1: General requirements (Class A). |

Signed for and on behalf of the above manufacturer

| | |
|----------------------------------|---|
| Additional Information | Year of CE Marking: 1996 |
| Place of Issue | Stone, Staffordshire, UK |
| Date of Issue | October 2009 |
| Authorised Representative | Carl Warren |
| Title | Technical Manager |
| Signature |  |

| Catalogue Number | Model | Description |
|------------------|---------------|--|
| 83055-02 | FF-200D-I-120 | PFP7, Industrial Flame Photometer, 110 V |
| 83055-05 | FF-200D-I | PFP7, Industrial Flame Photometer, 230 V |